

**Product Specification:**

Model no:	USI - LED 30D	USI - LED 30S
Intensity (in Lux)	160,000 + 160,000	160,000
No. of LED's	30+30	30
Light Field Diameter	150 mm	150 mm
LED service life	40,000 Hrs	40,000 Hrs
Dome Diameter	600 + 600 mm	600 mm
Colour Rendering Index (RA)	96	96
Temp. Increase @ Head Area	2 °C	2 °C
Control on Dome	Digital	Digital
Certificates	CE , WHO GMP, ISO 9001, 13485, 14001	
Input Voltage	AC 220 - 230V / 50 - 60 Hz	



Technology for life

**Corporate Office:**

445, F.I.E. Patparganj Industrial Area, Delhi-110092 (INDIA)  
Ph.: +91-11-22149600 / 94 / 95  
Fax: +91-11-43852679  
E-mail: info@unisurg.com  
Web.: www.unisurg.com

**Works:**

Pioneer Complex, 12/77-80, Site-IV, Sahibabad Industrial Area  
Distt. - Ghaziabad (U.P.) 201010




**United Surgical Industries**  
An ISO:9001/13485/14001 & CE Certified Company

Text and image file on this catalogue is property of United Surgical Industries, and protected by copyright and other restriction as well.  
© 2013, United Surgical Industries. All Rights reserved.

The company reserve the right to alter specification and design without our prior notice for the further improvement.

The new SUPERNOVA Series is based on LED Technology which offers glare free cold white light, akin to natural light which is ideal for modern surgeries. The lights incorporate international standards like CE mark, WHO GMP, ISO 9001, 13485, 14001 quality standards.



### Life of LED

LED service is 50 times longer than halogen bulb.



### Green & clean technology

LED lights are an extremely environmentally friendly alternative to conventional light sources. They contain no heavy metals, consume very little energy, and have virtually unlimited life.



### Shadow free light

Due to the multi-lens matrix with unique technology the light becomes shadowless on working area.

### Homogeneous Light Beam

The power from each light has been harnessed through a combination of finely honed lenses to deliver a homogeneous light.



### Focus Adjustment through Autoclavable handle

The handle for lamp holder may be pulled up and down, and used for turning lamp holder and adjusting focus. After Surgery, it may be removed easily and soaked in solution or disinfected under high pressure and high temperature.



### Digital Control Panel

All lights come with standard Control panel on each dome to adjust the Lux Intensity of the Light.



LED 30S

